## UTTAR BANGA KRISHI VISWAVIDYALAYA

## Pundibari, Cooch Behar, West Bengal- 736165

## **Inviting Quotation**

Tender Ref. No. UBKV/DR(T&P) -031, Date: 05.07.2024

Sealed quotations are invited form reputed suppliers for supplying of Binocular Microscope for the year2024 under "Bio-efficacy and phyto-toxicity of Boscalid 25.2% + Pyraclostrobin 12.8% WG against powdery mildew of Chilli" project, Uttar Banga Krishi Vishwavidyalaya . The details of the

SI. No	tions are herein giver	Specification	Quantity
01.	Binocular	Optical System: Semi-plan Achromat	
	Microscope	Body: Aluminum die cast body with all critical	01 pc.
		movements based on ball bearing and wire guides	
	F 1998 600 51-3	thereby ensuring smooth and precise manipulation	
	The Registralia	Binocular Observation Head: 45 deg with additional	
		safety lock facility	
	a Service Towns	Eyepiece: HWF 10X (FN 18 mm) paired eyepiece.	
	a Processorhund au	Nosepiece: Quadruple revolving nosepiece based on	
		precision	
		ball-bearing mechanism with positive stop click.	
		Objective: Semi Plan Achromat 4X, 1OX, 4OX (spring)	
	A. S.		
	and the second	and 100X oil spring	
	ge French 27 og	Mechanical Stage: Co-axial low drive mechanical stage	
	reges traingental ch	(125 mm X 145 mm) (+- 5mm) with traverse area of	
		50mm X 76mm (+ - 5 mm)	
		Condenser: Abbe condenser with aperture iris	
	ALLONG TO SELECT AND AND ADDRESS.	diaphragm (N.A 1.25) focusable with rack & oinion nd a	
	TO THE REPORT OF THE PARTY OF T	continuously variable iris diaphragm with a removable	
	mer and Higher the	blue filter for daylight	1000
	recommend to be a	observation. Rack & pinion mounted condenser holder.	
	Allegae Crischers C	Focusing: Co-axial coarse & fine controls with a focus	
		adjustment and fine adjustment knobs. Coarse focus	e signic a
		range 20 mm.	
		Illumination: Built-in illumination base with pre-	
		centered LED light sourse coupled with an efficient	
		collector lens system.	
		Universal power supply 100V-240V AC 50/60 HZ	
	9.30.0000000000000000000000000000000000	(SMPS). LED light should glow gradually at the switch	
	Designation of the second	it on at its maximum power.	
	L. Northead J	Microscope's base should be hexagonal design for	20.00
	1	better stability.	
	in the Subhern J	Power supply open from front side for easy remove of	
	011-15,900	light user friendly.	
		Should have 4-digital European CE certification on	
	Lamb.	model no.	

2 PSW 4 M	Manufacturer should have ISO 14001: 2015 for the	
LILAN	design, sales of microscopes & related services (Repair & OEM	
Page Barra	product development)	

Date of publication

: 05/07/2024

Last date of Submission

: 22/07/2024

up to 5 p.m.

## **Terms and Condition:**

- 1. Quotation to be submitted with a forwarding letter address to The Director of Research, UBKV, Pundibari, Cooch Behar, West Bengal
- 2. Covers of Quotations to be submitted to the office of the Director of Research's UBKV, Pundibari, Cooch Behar, West Bengal
- 3. Office hour with an email ID for further communication.
- 4. The technical Bid shall comprise of self attested photocopy of the following documents:
  - a. Valid copy of trade license.
  - b. GST Registration Certificate
  - c. Registration Certificate of P-Tax, along-with latest clearance certificate,
  - d. Service Tax Registration No; along with latest clearance certificate
  - e. Dealership/manufacturing/SSI certificate
  - f. ISO Certificate (if any)
  - g. Credentialed/user list with photograph,
  - h. Full Specification of items if applicable
  - i. Photocopy of PAN Card
  - 5. The Financial Bid Comprises of rate of items per unit including applicable taxes, Installation charges, incidental charges, extent of discount (as applicable)
  - 6. The Technical Bid and financial Bid should be submitted in two separate sealed envelopes with broad heading mentioning the name of the item and Tender No. on the top of the envelopes. The sealed envelope of Technical Bid and Financial Bid shall be submitted in a large sealed cover and further mentioning the name of the item and Tender number on the Top of the Envelope and to be addressed to the Director of Research, UBKV, Pundibari, Cooch Behar.
  - 7. The materials should be delivered by the Supplier at their own cost and responsibility.
  - 8. The University reserves the right to accept or reject any or all tenders without giving any reason whatsoever.

Sd/

Director of Research(Actg.)

Date: 05.07.2024

Ref. No. UBKV/DR(T&P)- 031 / (4)

Copy forwarded for information and necessary action to:

- 01. Notice Board, UBKV, Pundibari
- 02. Central Store Section, UBKV, Pundibari
- 03. Mr. Subhendu Das; He is requested to add in the University Website at the earliest
- 04. Office copy Director of Research, UBKV, Pundibari

Director of Research(Actg.)

Director of Research
Ulter Benga Krishi Viswavidyalaya
Pundibari, Cooch Behar